

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument  
State of Alaska

OCT 28 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130173</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Josh Hemburg</u>	ID# <u>5828</u> Date <u>10-25-09</u>
A Agency <u>AST</u>	Phone # <u>822-3263</u>
Instrument Location <u>Glennallen</u>	
Alco S/N <u>X124625</u> Target Value <u>0.083</u> High Pressure <u>1100</u>	
B Alco Test Values	<u>0.077</u> <u>0.077</u> <u>0.077</u>
1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature <u>[Signature]</u>	
(OVER)	

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 28<sup>th</sup> day of Dec, 2009.

[Signature] (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



State of Alaska  
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

(CONTINUED FROM FRONT PAGE)

C  
TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC  
CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

REQUESTED FOR 27.69 in	
.076	13:04
.000	13:06
VERIFIED	13:06
.077	13:06
.000	13:07
.000	13:07
.000	13:08
.078	13:09
.000	13:09

## DIAGNOSTIC CHECK

```

PRINTER TEST
"0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefg
hijklmno
pqrstuvwxyz{|}~

```